



Australian Government  
Department of Industry,  
Science and Resources

National  
Measurement  
Institute

# Proficiency Test Final Report AQA 23-13 Heroin

December 2023

© Commonwealth of Australia 2023.

Unless otherwise noted, the Commonwealth owns the copyright (and any other intellectual property rights, if any) in this publication.

All material in this publication is provided under a Creative Commons Attribution 4.0 International Licence (CC BY 4.0), with the exception of:

- the Commonwealth Coat of Arms;
- the logo of the Department of Industry, Science and Resources;
- photographs of our staff and premises; and
- content supplied by third parties.

Creative Commons Attribution 4.0 International Licence is a standard form licence agreement that allows you to copy, distribute, transmit and adapt this publication provided you attribute the work. A summary of the licence terms is available at:

[creativecommons.org/licenses/by/4.0/](https://creativecommons.org/licenses/by/4.0/). Further details are available on the Creative Commons website, at: [creativecommons.org/licenses/by/4.0/legalcode](https://creativecommons.org/licenses/by/4.0/legalcode).

You may not copy, distribute, transmit or adapt any material in this publication in any way that suggests that this department or the Commonwealth endorses you or any of your services or products.

### **Attribution**

Material contained in this publication is to be attributed to this department as:

© Commonwealth of Australia, Department of Industry, Science and Resources, Proficiency Test Final Report AQA 23-13 Heroin, 2023.

### **Third party copyright**

Wherever a third party holds copyright in material contained in this publication, the copyright remains with that party. Their permission may be required to use the material.

This department has made all reasonable efforts to:

- clearly label material where the copyright is owned by a third party;
- ensure that the copyright owner has consented to this material being contained in this publication.

### **Using the Commonwealth Coat of Arms**

The terms of use for the Coat of Arms are available on the Department of Prime Minister and Cabinet's website, at [www.pmc.gov.au/resource-centre/government/commonwealth-coat-arms-information-and-guidelines](http://www.pmc.gov.au/resource-centre/government/commonwealth-coat-arms-information-and-guidelines)

## ACKNOWLEDGMENTS

This study was conducted by the National Measurement Institute (NMI). Support funding was provided by the Australian Government Department of Industry, Science and Resources.

I would like to thank the management and staff of the participating laboratories for supporting the study. It is only through widespread participation that we can provide an effective service to laboratories.

The assistance of the following NMI staff members in the planning, conduct and reporting of the study is acknowledged.

Jenny Xu

Lesley Johnston

Raluca Iavetz

Manager, Chemical Reference Values

105 Delhi Rd, North Ryde, NSW 2113, Australia

Phone: +61 2 9449 0178

Email: [raluca.iavetz@measurement.gov.au](mailto:raluca.iavetz@measurement.gov.au)



Accredited for compliance with ISO/IEC 17043

## TABLE OF CONTENTS

SUMMARY	1
1 INTRODUCTION	2
1.1 NMI Proficiency Testing Program	2
1.2 Study Aims	2
1.3 Study Conduct	2
2 STUDY INFORMATION	3
2.1 Study Timetable	3
2.2 Participation and Laboratory Code	3
2.3 Test Material Specification	3
2.4 Test Sample Homogeneity and Stability	3
2.5 Sample Dispatch and Receipt	4
2.6 Instructions to Participants	4
2.7 Interim Report and Preliminary Report	4
3 PARTICIPANT LABORATORY INFORMATION	5
3.1 Test Methods Reported by Participants	5
3.2 Details of Participant Calibration Standards	6
3.3 Reported Basis of Participants' Measurement Uncertainty Estimates	8
3.4 Participants' Comments	10
4 PRESENTATION OF RESULTS AND STATISTICAL ANALYSIS	12
4.1 Results Summary	12
4.2 Outliers and Extreme Outliers	12
4.3 Assigned Value	12
4.4 Robust Average and Robust Standard Deviation	12
4.5 Performance Coefficient of Variation (PCV)	12
4.6 Target Standard Deviation for Proficiency Assessment	13
4.7 z-Score	13
4.8 $E_r$ -Score	13
4.9 Traceability and Measurement Uncertainty	13
5 TABLES AND FIGURES	14
6 DISCUSSION OF RESULTS	21
6.1 Assigned Value	21
6.2 Measurement Uncertainty Reported by Participants	21
6.3 z-Score	21
6.4 $E_r$ -Score	22
6.5 Identification of Cutting Agents	23
6.6 Participants' Analytical Methods	24
6.7 Comparison of Results and Date of Analysis	26
6.8 Comparison with Previous Heroin PT Studies	27
7 REFERENCES	30
APPENDIX 1 ROBUST AVERAGE AND ASSOCIATED UNCERTAINTY, z-SCORE AND $E_r$ -SCORE CALCULATIONS	31
APPENDIX 2 ACRONYMS AND ABBREVIATIONS	32

## SUMMARY

AQA 23-13 Heroin commenced in May 2023. Sets of heroin hydrochloride, each containing three test samples, were sent to 33 laboratories, with one laboratory requesting two sample sets to be analysed independently by different analysts. All participants submitted results. Samples were prepared at the Sydney NMI laboratory using heroin hydrochloride samples supplied by the Australian Federal Police.

The assigned values in this study were the robust averages of participants' results.

**Traceability:** The consensus of participants' results is not traceable to any external reference, so although expressed in SI units, metrological traceability has not been established.

The outcomes of the study were assessed against the aims as follows:

- *Assess the proficiency of participants measuring heroin in samples typical of a routine seizure.*

Participant performance was assessed by  $z$ -scores and  $E_n$ -scores.

Of 102  $z$ -scores, 86 (84%) returned  $|z| \leq 2.0$ , indicating an acceptable performance.

Of 102  $E_n$ -scores, 93 (91%) returned  $|E_n| \leq 1.0$ , indicating agreement of the participant's result with the assigned value within their respective expanded uncertainties.

Laboratories **1, 3, 6, 9, 10, 11, 12, 14, 15, 16, 17, 18, 19, 20, 23, 25, 26, 27, 28, 29, 30, 33** and **34** returned acceptable  $z$ -scores and  $E_n$ -scores for all results.

- *Develop a practical application of traceability and measurement uncertainty, and provide participants with information that will assist uncertainty estimates.*

Of 102 reported results, 96 (94%) were reported with an associated expanded measurement uncertainty. The magnitude of reported uncertainties was within the range 2.9% to 26% relative.

- *Test the ability of participants to identify cutting agents commonly found in controlled drug preparation.*

Sample S1 was cut with niacinamide, Sample S2 was cut with sucrose, and Sample S3 was cut with procaine hydrochloride.

All participants reported on the identity of at least one sample's cutting agent. Laboratories **3, 4, 7, 9, 10, 16, 17, 22, 25, 29, 30** and **31** correctly reported all cutting agents used.

- *Produce materials that can be used in method validation and as control samples.*

The test samples produced for this study are homogeneous and well characterised. Surplus samples are available for purchase and can be used for quality control and for method validation purposes.

## **1 INTRODUCTION**

### **1.1 NMI Proficiency Testing Program**

The National Measurement Institute (NMI) is responsible for Australia's national measurement infrastructure, providing a range of services including a chemical proficiency testing program.

Proficiency testing (PT) is the 'evaluation of participant performance against pre-established criteria by means of interlaboratory comparison'.<sup>1</sup> NMI PT studies target chemical testing in areas of high public significance such as trade, environment, law enforcement and food safety. NMI offers studies in:

- pesticide residues in fruit, vegetables and herbs, soil and water;
- petroleum hydrocarbons in soil and water;
- per- and polyfluoroalkyl substances in water, soil, biota and food;
- inorganic analytes in soil, water, filters, food and pharmaceuticals;
- controlled drug assay, drugs in wipes and clandestine laboratory; and
- allergens in food.

### **1.2 Study Aims**

The aims of the study were to:

- assess the proficiency of participants measuring heroin in samples typical of a routine seizure;
- develop a practical application of traceability and measurement uncertainty, and provide participants with information that will assist uncertainty estimates;
- test the ability of participants to identify cutting agents commonly found in controlled drug preparation; and
- produce materials that can be used in method validation and as control samples.

The choice of the test method was left to the participating laboratories.

### **1.3 Study Conduct**

The conduct of NMI proficiency tests is described in the NMI Study Protocol for Proficiency Testing.<sup>2</sup> The statistical methods used are described in the NMI Chemical Proficiency Testing Statistical Manual.<sup>3</sup> These documents have been prepared with reference to ISO/IEC 17043,<sup>1</sup> and The International Harmonized Protocol for the Proficiency Testing of Analytical Chemistry Laboratories.<sup>4</sup>

NMI is accredited by the National Association of Testing Authorities, Australia (NATA) to ISO/IEC 17043 as a provider of proficiency testing schemes.<sup>1</sup> This controlled drug study is within the scope of NMI's accreditation.

## **2 STUDY INFORMATION**

### **2.1 Study Timetable**

The timetable of the study was:

Invitations sent	22/05/2023
Samples sent	27/07/2023
Results due	3/11/2023
Interim Report	8/11/2023
Preliminary Report	8/11/2023

The study timeline was extended to accommodate sample delivery delays to some international participants.

### **2.2 Participation and Laboratory Code**

Thirty-three laboratories registered to participate, with one laboratory requesting two sets of samples each to be analysed independently by different analysts. All participants were assigned a confidential laboratory code number for this study. All participants submitted results.

### **2.3 Test Material Specification**

Three test samples were prepared in July 2023. The starting material was heroin hydrochloride (approximately 80% heroin base (m/m)) supplied by the Australian Federal Police.

Niacinamide (nicotinamide), sucrose and procaine hydrochloride purchased from Sigma-Aldrich were used as cutting agents. Sample S1 was cut with niacinamide, Sample S2 was cut with sucrose, and Sample S3 was cut with procaine hydrochloride.

The heroin hydrochloride was ground and sieved through a 180 µm sieve. The cutting agents were processed similarly. Test samples were prepared by adding a known mass of sieved cutting agent to a known mass of sieved drug material, and then mixing this in a tumbler overnight. Portions of 150 mg of each of the test samples were weighed into labelled glass vials.

Sample S1 was prepared to contain approximately 15% heroin base (m/m).

Sample S2 was prepared to contain approximately 30% heroin base (m/m).

Sample S3 was prepared to contain approximately 66% heroin base (m/m).

### **2.4 Test Sample Homogeneity and Stability**

The preparation of homogeneous test samples is an important part of a PT study. Given the small (usually < 150 mg) test portions normally used for controlled substances analysis, the particle size must be sufficiently small and uniformly distributed to ensure minimal influence on analytical precision. The procedure for the preparation of the study samples has been validated in previous studies, and no additional homogeneity testing was conducted in this proficiency study. Results returned by the participants also gave no reason to question the homogeneity of the test samples.

To assess stability of the samples, results returned by participants were compared to the date of analysis. The results gave no reason to question the stability of the test samples (Section 6.7).

## 2.5 Sample Dispatch and Receipt

Sets of three test samples, with each sample containing approximately 150 mg of material, were dispatched to participants on 27 July 2023.

The following items were also sent with the samples:

- a covering letter which included a description of the test samples and instructions for participants; and
- a form for participants to confirm the receipt of the test samples.

An Excel spreadsheet for the electronic reporting of results was emailed to participants.

## 2.6 Instructions to Participants

Participants were instructed as follows:

- Analyse each sample for amount of heroin base by your routine test method.
- Identify and report the diluent(s) and/or adulterant(s) in all samples if this is within your normal scope of analysis.
- For each sample, report % m/m heroin as base. Report this figure as if reporting to a client.
- For each result, report an estimate of your expanded uncertainty as % m/m heroin as base.
- Give brief details of your:
  - basis of uncertainty estimate (e.g. uncertainty budget, repeatability precision)
  - analytical method (e.g. sample treatment, instrument type, calibration method)
  - reference standard (e.g. source, purity)as requested by the results sheet.
- Please complete the results sheet by Wednesday 20 September 2023 and return by email to [jenny.xu@measurement.gov.au](mailto:jenny.xu@measurement.gov.au). Late results may not be included in the study report.

The results due date was extended to 3 November 2023 for all participants. This was due to significant sample delivery delays to several international participants, caused by delays with receiving some export permits.

## 2.7 Interim Report and Preliminary Report

An Interim Report was emailed to all participants on 8 November 2023.

A Preliminary Report was emailed to all participants on 8 November 2023. This report included a summary of the results reported by participants, assigned values, performance coefficients of variation (PCVs),  $z$ -scores and  $E_n$ -scores for each analyte in this study. No data from the Preliminary Report has been changed in the present Final Report.



### 3 PARTICIPANT LABORATORY INFORMATION

#### 3.1 Test Methods Reported by Participants

Participants' reported test methods are presented in Table 1. Responses may have been modified so that the participant cannot be identified.

Table 1 Summary of Participants' Test Methods

Lab. Code	Extraction Solvent	Internal Standard	Calib. Points	Technique	Detector	Column
1	Chloroform/Methanol	2,2,2-Triphenylacetophenone	1	GC	FID	HP5
2	Ethanol	Propylparaben	8	UPLC	DAD	BEH Shield RP18
3	Acetonitrile/Water	None	5	HPLC	UV	Kinetex 5u C18
4	water/acetonitrile/2.5M sulphuric acid 90:10:1	None	3	HPLC	Diode Array	Shimpack XR-ODS
5	Methanol	none	2	HPLC	DAD	Luna 3 µm PFP 100 Å 150x4.6 mm
6	Chloroform:methanol (9:1)	Triphenylacetophenone	1	GC	FID	HP5
7	Acetonitrile	Strychnine	6	GC	FID	Phenyl
8	Chloroform	octacosane	5	GC	MS	Rxi-5Sil-MS
9	Water:Acetonitrile	-	3	HPLC	UV/Vis	Luna C18
10	acetonitrile/H2O (80/20)	External standard	3	HPLC	DAD	NH2
11	methanol	NO	7	HPLC	DAD	Poroshell 120 C18 (4.6X150mm, 2.7 microns particle size)
12	Chloroform	Nortriptyline	1	GC	FID	HP5
13	ethanol:dimethylformamide (9:1)	tribenzylamine	6	GC	FID	HP1
14	acetonitrile / water	none	1	HPLC	UV/Vis	Kromasil
15	Acetonitrile:Water (75:25)	Benzocaine	3	UPLC	DAD	Acquity UPLC BEH C18 1.7um (2.1x100mm)
16	Chloroform	2,2,2-triphenylacetophenone	S1, S2: 1 S3: 4	GC	FID	HP-1

Lab. Code	Extraction Solvent	Internal Standard	Calib. Points	Technique	Detector	Column
17	Acetonitrile, acetic acid, water	NO ISTD	4	HPLC	UV DAD	Poroshell 120 Ec-18
18	Methanol	N/A	3	HPLC	PDA	Silica 15cm
19	Chloroform	Octacosane	5	GC	FID	HP5
20	Chloroform	Nortriptyline	1	GC	FID	HP5
21	ACN/MeOH/H2O	Analog of heroin	7	UPLC	MS/MS	C-18 column
22	Ethanol absolute	Tribenzylamine	6	GC	FID	DB5
23	Methanol	Diazepam	6	GC	FID	J&W 128-5512
24	METHANOL	LOXAPINE	5	HPLC	DAD	XTERRA
25	Acetonitrile	None	7	HPLC	UV/Vis	Luna 3µm C8(2) 100Å, 100x2mm
26	Acetonitrile/Methanol (95:5)	Pholcodine 1mg/ml	3	UPLC	PDA	ACQUITY C-18
27	Methanol	Mepivacaine	4	UPLC	DAD	Kinetex Evo C18
28	Methanol	none	5	HPLC	DAD	Kinetex C-18-XB
29	Ethanol	Triphenylacetophenone (TPAP)	3	GC	FID	HP1-MS
30	Methanol	Methadone	4	GC	FID	Rxi-5ms
31	acetonitrile/water (86/14)	none	4	HPLC	DAD	NH2
32	Ethanol	Eicosane	6	GC	FID	HP5
33	HPLC Methanol	-	1	UPLC	DAD	Thermo Scientific Hypersil-5-ODS
34	HPLC Methanol	-	1	UPLC	DAD	Thermo Scientific Hypersil-5-ODS

### 3.2 Details of Participant Calibration Standards

Participants' responses regarding their calibration standard are presented in Table 2. Responses may have been modified so that the participant cannot be identified.

Table 2 Participant Calibration Standard

Lab. Code	Reference Standard	Purity (%)
1	Toronto Research Chemicals	98
2	NMI	99.4
3	British Pharmacopoeia	99.3
4	LGC (Mikromol)	99.7
5	Lipomed (M-29-FB-1LA)	1 mg/mL
6	in-house reference material	98
7	NMI	99.3
8	Lipomed	99.912
9	British Pharmacopoeia	99.3
10	NMI	99.4
11	Lipomed	99.879
12	Inhouse Reference Material	98
13	Lipomed	99.912
14	Lipomed	99.91
15	NMI	99.3
16	In-house synthesis	97.3
17	Lipomed	99.88
18	Johnson Matthey	99.4
19	NMI	99.3
20	Inhouse Reference Material	99.8
21	Lipomed	100
22	Lipomed	99.912% +/- 0.018% free base content 86.4%
23	Lipomed	99.1
24	LGC STANDARDS	>99.9%
25	NMI	99.4
26	NMI	99.3
27	Lipomed	99.600 +/- 0.020
28	Chiron	99.8 (±4.0)
29	NMI	99.3 +/- 1.3
30	LGC	1.011mg/ml
31	Lipomed	99.912
32	Alcaliber	98.4
33	Lipomed	99.912 ± 0.018
34	Lipomed	99.912 ± 0.018

### 3.3 Reported Basis of Participants' Measurement Uncertainty Estimates

Participants were requested to provide information about their basis of measurement uncertainty (MU). Responses are presented in Table 3. Responses may have been modified so that the participant cannot be identified.

Table 3 Reported Basis of Uncertainty Estimate

Lab. Code	Approach to Estimating MU	Information Sources for MU Estimation*		Guide Document for Estimating MU
		Precision	Method Bias	
1	Top Down - precision and estimates of the method and laboratory bias	Duplicate analysis	Homogeneity of sample Masses and volumes Matrix effects Standard purity	ASCLD/LAB Guidance On The Estimation Of Measurement Uncertainty, AL-PD-3061
2				
3	Top Down - precision and estimates of the method and laboratory bias	Control samples Duplicate analysis	Instrument calibration Homogeneity of sample Masses and volumes Matrix effects	Eurachem/CITAC Guide
4	Standard deviation of replicate analyses multiplied by 2 or 3	Control samples - RM Duplicate analysis	Instrument calibration Standard purity	ISO/GUM
5	Top Down - precision and estimates of the method and laboratory bias	Control samples - Samples from case Duplicate analysis	Laboratory bias from PT studies	Nordtest Report TR537
6	Top Down - precision and estimates of the method and laboratory bias	Control samples - RM	Homogeneity of sample Matrix effects Standard purity	ASCLD/LAB Guidance on the Estimation of Measurement Uncertainty, AL-PD-3061
7	Top Down - precision and estimates of the method and laboratory bias	Control samples - CRM Duplicate analysis	Instrument calibration Homogeneity of sample Masses and volumes Laboratory bias from PT studies	ISO/GUM
8	Top Down - precision and estimates of the method and laboratory bias	Control samples	Laboratory bias from PT studies	ISO/GUM
9	Precision and Bias	Control samples - Known Value Samples	Instrument calibration Homogeneity of sample Masses and volumes Recoveries of SS	ISO/GUM
10	Top Down - precision and estimates of the method and laboratory bias	Duplicate analysis	Instrument calibration Standard purity	ISO/GUM
11	Top Down - precision and estimates of the method and laboratory bias	Control samples - CRM Duplicate analysis	Instrument calibration Homogeneity of sample Masses and volumes Laboratory bias from PT studies Standard purity	

Lab. Code	Approach to Estimating MU	Information Sources for MU Estimation*		Guide Document for Estimating MU
		Precision	Method Bias	
12	Top Down - precision and estimates of the method and laboratory bias	Duplicate analysis	Instrument calibration	ISO/GUM
13	Top Down - precision and estimates of the method and laboratory bias	Control samples - RM	Standard purity	
14	Standard deviation of replicate analyses multiplied by 2 or 3	Control samples - RM		ISO/GUM
15	Top Down - precision and estimates of the method and laboratory bias	Control samples - RM Duplicate analysis	Masses and volumes Standard purity	Eurachem/CITAC Guide
16	Validation			
17	Top Down - precision and estimates of the method and laboratory bias			ISO 5725-2 and ISO/TS 21748
18	Standard deviation of replicate analyses multiplied by 2 or 3	Control samples - SS	Instrument calibration Masses and volumes Recoveries of SS	ISO/GUM
19	Top Down - precision and estimates of the method and laboratory bias	Control samples - previously analysed police seizures Duplicate analysis	Instrument calibration Homogeneity of sample Masses and volumes Matrix effects Standard purity	Eurachem/CITAC Guide
20	Top Down - precision and estimates of the method and laboratory bias	Duplicate analysis	Masses and volumes Recoveries of SS	Eurachem/CITAC Guide
21	Top Down - precision and estimates of the method and laboratory bias			
22	Black Box	Control samples - CRM Duplicate analysis	Instrument calibration Homogeneity of sample Masses and volumes Matrix effects Laboratory bias from PT studies Standard purity	Eurachem/CITAC Guide
23	Estimating Measurement Uncertainty by black box with pairs of values	Standard deviation from PT studies only		ISO/GUM ENAC G 09 or ISO 21748
24				
25	Top Down - precision and estimates of the method and laboratory bias	Control samples - In House Control Duplicate analysis	Instrument calibration Homogeneity of sample Masses and volumes Standard purity	ISO/GUM
26	Top Down - precision and estimates of the method and laboratory bias	Control samples - CRM Duplicate analysis	Instrument calibration Homogeneity of sample Masses and volumes	NATA GAG Estimating and Reporting Measurement

Lab. Code	Approach to Estimating MU	Information Sources for MU Estimation*		Guide Document for Estimating MU
		Precision	Method Bias	
				Uncertainty of Chemical Test Results
27	Top Down - precision and estimates of the method and laboratory bias	Control samples Duplicate analysis	Instrument calibration Homogeneity of sample Matrix effects Laboratory bias from PT studies Recoveries of SS Standard purity	EA-04/16 EA guidelines on the expression of uncertainty in quantitative testing.
28	Standard deviation of replicate analyses multiplied by 2 or 3	Control samples - RM Duplicate analysis		Eurachem/CITAC Guide
29	Bottom Up (ISO/GUM, fish bone/cause and effect diagram)	Control samples - CRM Duplicate analysis	Instrument calibration Homogeneity of sample Masses and volumes Matrix effects Recoveries of SS Standard purity	Eurachem/CITAC Guide
30	Standard deviation of replicate analysis multiplied by 2 or 3	Duplicate analysis	Masses and volumes	ISO/GUM
31	Top Down - precision and estimates of the method and laboratory bias	Control samples - RM	Laboratory bias from PT studies	NF ISO 11352 and NF V03-110
32	Top Down - precision and estimates of the method and laboratory bias	Control samples - RM Duplicate analysis	Matrix effects	ISO/GUM
33	Top Down - precision and estimates of the method and laboratory bias	Control samples - SS Duplicate analysis	Laboratory bias from PT studies Recoveries of SS	Eurachem/CITAC Guide
34	Top Down - precision and estimates of the method and laboratory bias	Control samples - SS Duplicate analysis	Laboratory bias from PT studies Recoveries of SS	Eurachem/CITAC Guide

\*SS = Spiked Samples, RM = Reference Material, CRM = Certified Reference Material

### 3.4 Participants' Comments

Participants were invited to comment on the samples, their methodology, the PT study in general and suggestions for future PT studies. Such feedback allows for the improvement of future studies. Participants' comments are presented in Table 4, along with the study coordinator's response where appropriate. Some responses may be modified so that the participant cannot be identified.

Table 4 Participant Comments

Lab. Code	Participants' Comments	Study Coordinator's Response
1	Methodology: Quantitation of heroin and monoacetylmorphines by GC-FID	
4	Uncertainty: MuM determined from multiple injections of reference material. $3 \times (\text{Std Dev}/\text{mean}) \times 100$ .	

Lab. Code	Participants' Comments	Study Coordinator's Response
6	Methodology: Quantitation for heroin and monoacetylmorphines by GC	
9	Methodology: The sample is weighed and extracted in duplicate and the components separated by HPLC with UV detection. The percentage purity is then determined by comparison with a standard calibration curve.	
11	Methodology: External standard	
15	Heroin and acetylcodeine detected in S1, S2 & S3.	
17	Methodology: 0 ; 5 ; 20 ; 100 mg/l	
18	Routine case samples would always be round down i.e -3.92% for example $15.3 \times 0.9608 = 14$ Uncertainty: The reported result (in routine case samples) is defined as the mean of the individual results multiplied by the uncertainty correction factor and is rounded down to the nearest whole number (unless <1% w/w). E.g a mean result of 15.3% with an uncertainty correction value of 96.08% would give a reported result of $15.3 \times 0.9608 = 14.7$ therefore rounded down to 14%.	
20	Methodology: Quantification of heroin by GC-FID	
22	Methodology: Dilution of sample in 10 mL of iSTD (0.25 mg/mL of TBA in abs. ETOH)	
25	It is acknowledged that Acetylcodeine and Monoacetyl morphine are not diluents/adulterants but it is standard laboratory practice to report these. Niacinamide was tentatively identified based on comparison of the compound's mass spectrum and FTIR spectrum with that of a literature source. Unequivocal identification of this compound can be made on receipt of a certified reference material at the laboratory.	
29	Methodology: Dichloromethane (30ml/L of ethanol) was used to dissolve the TPAP	
30	Could the sample vials be submitted without the name of the analyte being present - this would enable a full blind test by the analyst rather than them being prompted straight away as to what the sample is.	This PT is not a qualitative study. All participants are informed what analyte they are assessing for, on the sample label as well as the sample dispatch letter provided with the samples.
31	we would like to receive 3 samples of very different concentration for example 3%, 30% and 80% Methodology: Eluant acetonitrile/water (86/14) + 2.25ml picA/litre	A range of drug purities are selected to cater for the needs of different laboratories. In this study, the samples contained 14.8%, 29.3% and 65.3% heroin base (m/m).

## 4 PRESENTATION OF RESULTS AND STATISTICAL ANALYSIS

### 4.1 Results Summary

Participant results are listed in Tables 5 to 7 with summary statistics: robust average, median, mean, number of numerical results (N), maximum (Max), minimum (Min), robust standard deviation (Robust SD) and robust coefficient of variation (Robust CV).

Bar charts of results and performance scores are presented in Figures 2 to 4. An example chart with interpretation guide is shown in Figure 1.




Figure 1 Guide to Presentation of Results

### 4.2 Outliers and Extreme Outliers

Outliers were results less than 50% and greater than 150% of the robust average, and these were removed before the calculation of the assigned value, if applicable.<sup>3,4</sup> Extreme outliers were obvious blunders, e.g. results reported with incorrect units or for a different analyte or sample, and such results were removed for the calculation of all summary statistics.<sup>3,4</sup>

### 4.3 Assigned Value

The assigned value is defined as the 'value attributed to a particular property of a proficiency test item'.<sup>1</sup> In this study, the property is the % heroin base (m/m) in the test samples. Assigned values were the robust averages of participants' results and the expanded uncertainties were estimated from the associated robust SDs (Appendix 1).

### 4.4 Robust Average and Robust Standard Deviation

The robust averages and associated expanded MUs, and robust SDs (a measure of the variability of participants' results), were calculated using the procedure described in ISO 13528.<sup>5</sup>

### 4.5 Performance Coefficient of Variation (PCV)

The PCV is a fixed measure of the between-laboratory variation that in the judgement of the study coordinator would be expected from participants given the levels of analytes present, and is supported by mathematical models such as the Thompson-Horwitz equation.<sup>6</sup> It is important to note that this is a performance measure set by the study coordinator and it is not the robust CV of participants' results. By setting a fixed and realistic value for the PCV, a participant's performance does not depend on other participants' performances, and can be compared from study to study.



#### 4.6 Target Standard Deviation for Proficiency Assessment

The target standard deviation for proficiency assessment ( $\sigma$ ) is the product of the assigned value ( $X$ ) and the PCV, as presented in Equation 1.

$$\sigma = X \times PCV \quad \text{Equation 1}$$

#### 4.7 z-Score

For each participant's result, a z-score is calculated according to Equation 2.

$$z = \frac{(\chi - X)}{\sigma} \quad \text{Equation 2}$$

where:

$z$  is z-score

$\chi$  is a participant's result

$X$  is the assigned value

$\sigma$  is the target standard deviation for proficiency assessment from Equation 1

For the absolute value of a z-score:

- $|z| \leq 2.0$  is acceptable;
- $2.0 < |z| < 3.0$  is questionable; and
- $|z| \geq 3.0$  is unacceptable.

#### 4.8 $E_n$ -Score

The  $E_n$ -score is complementary to the z-score in assessment of laboratory performance.  $E_n$ -score includes measurement uncertainty and is calculated according to Equation 3.

$$E_n = \frac{(\chi - X)}{\sqrt{U_\chi^2 + U_X^2}} \quad \text{Equation 3}$$

where:

$E_n$  is  $E_n$ -score

$\chi$  is a participant's result

$X$  is the assigned value

$U_\chi$  is the expanded uncertainty of the participant's result

$U_X$  is the expanded uncertainty of the assigned value

For the absolute value of an  $E_n$ -score:

- $|E_n| \leq 1.0$  is acceptable; and
- $|E_n| > 1.0$  is unacceptable.

#### 4.9 Traceability and Measurement Uncertainty

Laboratories accredited to ISO/IEC 17025 must establish and demonstrate the traceability and measurement uncertainty associated with their test results.<sup>7</sup>

Guidelines for quantifying uncertainty in analytical measurement are described in the Eurachem/CITAC Guide.<sup>8</sup>

## 5 TABLES AND FIGURES

Table 5

### Sample Details

<b>Sample No.</b>	S1
<b>Matrix</b>	Powder
<b>Analyte</b>	Heroin
<b>Unit</b>	% base (m/m)

### Participant Results

Lab. Code	Result	Uncertainty	z	E <sub>n</sub>
1	14.6	1.5	-0.45	-0.13
2	15.3	NR	1.13	1.67
3	14.60	2.04	-0.45	-0.10
4	15.87	1.18	2.41	0.88
5	13.7	1.1	-2.48	-0.96
6	14.1	1.7	-1.58	-0.41
7*	23.7	1	20.05	8.52
8	16.9	3.4	4.73	0.62
9	14.1	0.8	-1.58	-0.82
10	15.3	2.3	1.13	0.22
11	14	1	-1.80	-0.77
12	15	1.3	0.45	0.15
13	13.87	1.9	-2.09	-0.48
14	15.3	3.1	1.13	0.16
15	14.6	1.3	-0.45	-0.15
16	14.2	2.0	-1.35	-0.30
17	14.55	1.03	-0.56	-0.23
18	15.3	3.92	1.13	0.13
19	14.1	1.0	-1.58	-0.67
20	14.9	1.3	0.23	0.07
21	14	2.1	-1.80	-0.38
22	16.0	1.2	2.70	0.97
23	14.8	0.9	0.00	0.00
24	15.9	NR	2.48	3.67
25	14.7	3	-0.23	-0.03
26	14.1	1.7	-1.58	-0.41
27	14.9	1.0	0.23	0.10
28	15	0.8	0.45	0.23
29	14.4	0.7	-0.90	-0.53
30	14.21	0.94	-1.33	-0.60
31	18.39	1.66	8.09	2.13
32	15.8	1.6	2.25	0.61
33	15.0	0.9	0.45	0.21
34	15.3	1.0	1.13	0.48

\* Outlier, see Section 4.2

### Statistics

<b>Assigned Value</b>	14.8	0.3
<b>Robust Average</b>	14.9	0.4
<b>Median</b>	14.9	0.3
<b>Mean</b>	15.2	
<b>N</b>	34	
<b>Max</b>	23.7	
<b>Min</b>	13.7	
<b>Robust SD</b>	0.83	
<b>Robust CV</b>	5.6%	






Figure 2

Table 6

## Sample Details

<b>Sample No.</b>	S2
<b>Matrix</b>	Powder
<b>Analyte</b>	Heroin
<b>Unit</b>	% base (m/m)

## Participant Results

Lab. Code	Result	Uncertainty	z	E <sub>n</sub>
1	29.1	3.1	-0.23	-0.06
2	29.9	NR	0.68	1.50
3	28.22	3.95	-1.23	-0.27
4	29.92	2.22	0.71	0.27
5	28.5	2.3	-0.91	-0.34
6	29.8	3.7	0.57	0.13
7	28.9	1.3	-0.46	-0.29
8	35.0	7.0	6.48	0.81
9	29.3	1.6	0.00	0.00
10	29.4	4.5	0.11	0.02
11	29.2	1.4	-0.11	-0.07
12	29.6	2.7	0.34	0.11
13	28.3	1.4	-1.14	-0.69
14	28.4	5.7	-1.02	-0.16
15	29.1	2.1	-0.23	-0.09
16	28.8	4.2	-0.57	-0.12
17	29.64	1.98	0.39	0.17
18	30.1	3.92	0.91	0.20
19	28.6	2.0	-0.80	-0.34
20	28.2	2.5	-1.25	-0.43
21	27	3.8	-2.62	-0.60
22	31.2	2.3	2.16	0.81
23	29.8	1.8	0.57	0.27
24	34.5	NR	5.92	13.00
25	29.8	3	0.57	0.17
26	27.8	1.8	-1.71	-0.81
27	29.7	2.1	0.46	0.19
28	29	1.5	-0.34	-0.19
29	29.2	1.5	-0.11	-0.06
30	27.89	1.85	-1.60	-0.74
31	29.29	2.64	-0.01	0.00
32	30.3	3	1.14	0.33
33	30.2	1.9	1.02	0.46
34	30.2	1.9	1.02	0.46

## Statistics

<b>Assigned Value</b>	29.3	0.4
<b>Robust Average</b>	29.3	0.4
<b>Median</b>	29.3	0.4
<b>Mean</b>	29.5	
<b>N</b>	34	
<b>Max</b>	35	
<b>Min</b>	27	
<b>Robust SD</b>	0.95	
<b>Robust CV</b>	3.2%	






Figure 3

Table 7

**Sample Details**

<b>Sample No.</b>	S3
<b>Matrix</b>	Powder
<b>Analyte</b>	Heroin
<b>Unit</b>	% base (m/m)

**Participant Results**

Lab. Code	Result	Uncertainty	z	E <sub>n</sub>
1	65.4	6.9	0.05	0.01
2	67.5	NR	1.12	3.14
3	63.04	3.15	-1.15	-0.70
4	64.31	4.77	-0.51	-0.21
5	63.5	5.1	-0.92	-0.35
6	66.1	8.3	0.41	0.10
7*	25.8	1.1	-20.16	-30.30
8	78.5	15.7	6.74	0.84
9	64.2	3.5	-0.56	-0.31
10	67.2	10.1	0.97	0.19
11	65.7	2.7	0.20	0.14
12	65	5.9	-0.15	-0.05
13	63.54	3.2	-0.90	-0.54
14	65.8	13.2	0.26	0.04
15	63.0	4.6	-1.17	-0.49
16	65.0	1.9	-0.15	-0.15
17	65.44	3.46	0.07	0.04
18	65.7	3.92	0.20	0.10
19	63.1	4.4	-1.12	-0.49
20	66.7	6.1	0.71	0.23
21	60	8.4	-2.71	-0.63
22	67.2	5	0.97	0.38
23	66.2	4.0	0.46	0.22
24	64.4	NR	-0.46	-1.29
25	66.0	6.6	0.36	0.11
26	65.6	2.8	0.15	0.10
27	66.5	4.7	0.61	0.25
28	65	3.3	-0.15	-0.09
29	64.9	3.2	-0.20	-0.12
30	64.72	4.3	-0.30	-0.13
31	66.41	5.98	0.57	0.18
32	64.4	3.5	-0.46	-0.25
33	66.2	4.0	0.46	0.22
34	66.0	4.0	0.36	0.17

\* Outlier, see Section 4.2

**Statistics**

<b>Assigned Value</b>	65.3	0.7
<b>Robust Average</b>	65.2	0.7
<b>Median</b>	65.4	0.6
<b>Mean</b>	64.4	
<b>N</b>	34	
<b>Max</b>	78.5	
<b>Min</b>	25.8	
<b>Robust SD</b>	1.6	
<b>Robust CV</b>	2.4%	






Figure 4

Table 8 Reported Cutting Agents\*

Lab. Code	Cutting Agents		
	S1	S2	S3
Preparation	Niacinamide (nicotinamide)	Sucrose	Procaine hydrochloride
1	Niacinamide	-	Procaine
2	Nicotinamide		Procaine: 14.5 %
3	Niacinamide	Sucrose	Procaine, Acetyl codeine
4	Niacinamide	Sucrose	Procaine
5	Nicotinamide		Procaine
6	niacinamide	-	procaine
7	Niacinamide (Nicotinamide)	Sucrose	Procaine
8	acetylcodeine / niacinamide	acetylcodeine / 6-MAM	acetylcodeine / 6-MAM / procaine
9	Niacinamide	Sucrose	Procaine
10	Nicotinamide	Sucrose	Procaine
11	Nicotinamide		Procaine
12	Niacinamide	-	Procaine
13	Nicotinamide	Sugars	Procaine
14	Nicotinamide		Procaine
15	Nicotinamide indicated		Procaine indicated
16	Acetylcodeine, Nicotinamide	Acetylcodeine, Sucrose	Acetylcodeine, Procaine
17	Nicotinamide	Saccharose	Procaine
18	Niacinamide		Procaine
19	niacinamide		
20	Niacinamide	No	Procaine
21	none	none	procaine
22	6-MAM, Niacinamide, Acetylcodeine	6-MAM, Acetylcodeine, Sucrose	6-MAM, Acetylcodeine, Procaine
23	niacinamide, acetylcodeine, 6-monoacetylmorphine.	acetylcodeine, 6-monoacetylmorphine.	procaine, acetylcodeine, 6-monoacetylmorphine.
24	NICOTINAMIDE		PROCAINE
25	Acetylcodeine, Monoacetylmorphine, Niacinamide* (Tentative Identification)	Acetylcodeine, Monoacetylmorphine, Sucrose	Acetylcodeine, Monoacetylmorphine, Procaine
26	Nicotinamide		Procaine
27	Nicotinamide		Procaine
28	Nicotinamide, acetylcodeine	Acetylcodeine	Procaine, acetylcodeine
29	Nicotinamide	Sucrose	Procaine
30	Niacinamide	Sucrose	Procaine
31	nicotinamide	saccharose	procaine
32	Nicotinamide		Procaine
33	Nicotinamide	-	Procaine
34	Nicotinamide	-	Procaine

\* Responses may have been modified so that the participant cannot be identified.



## 6 DISCUSSION OF RESULTS

### 6.1 Assigned Value

The robust averages and associated expanded uncertainties were calculated using the procedure described in ISO 13528.<sup>5</sup> The assigned values for all scored analytes were the robust averages of participants' results, after results less than 50% and greater than 150% of the robust average had been removed.<sup>3,4</sup> The calculation of the expanded uncertainty for a robust average, using Sample S2 as an example, is presented in Appendix 1.

**Traceability:** The consensus of participants' results is not traceable to any external reference, so although expressed in SI units, metrological traceability has not been established.

### 6.2 Measurement Uncertainty Reported by Participants

Participants were asked to report an estimate of the expanded MU associated with their results and the basis of this uncertainty estimate (Table 3). One participant reported using the NATA GAG Estimating and Reporting MU as their guide; NATA no longer publishes this.<sup>9</sup>

It is a requirement of ISO/IEC 17025 that laboratories have procedures to estimate the uncertainty of chemical measurements and to report this uncertainty in specific circumstances, including when the client's instruction so requires.<sup>7</sup>

Of 102 reported results, 96 (94%) were reported with an associated expanded MU. Laboratories **2** and **24** did not report any uncertainties; these participants reported that they were not accredited.

The magnitude of reported uncertainties was within the range 2.9% to 26% relative. In general, an expanded uncertainty of less than 3% may be unrealistically small for the routine measurement of illicit drugs, while over 10% may be too large and not fit for purpose. Of the 96 expanded MUs, 66 (69%) were between 3% and 10% relative to the result, one was less than 3% and 29 were greater than 10%.

Uncertainties associated with results returning a satisfactory *z*-score but an unsatisfactory *E<sub>n</sub>*-score may have been underestimated.

In some cases, results were reported with an inappropriate number of significant figures. Including too many significant figures may inaccurately reflect measurement precision. The recommended format is to write the uncertainty to no more than two significant figures, and then to write the result with the corresponding number of decimal places. For example, instead of reporting  $65.44 \pm 3.46\%$ , the recommended format is  $65.4 \pm 3.5\%$ .<sup>8</sup>

### 6.3 z-Score

Target SDs equivalent to 3% PCV were used to calculate *z*-scores. CVs predicted by the Thompson-Horwitz equation,<sup>6</sup> target SDs (as PCVs) and between-laboratory CVs obtained in this study are presented for comparison in Table 9.

Table 9 Comparison of Thompson-Horwitz CVs, Between-Laboratory CVs and Target SDs

Sample	Analyte	Assigned Value (% base (m/m))	Thompson-Horwitz CV <sup>a</sup> (%)	Between-Laboratory CV <sup>b</sup> (%)	Target SD (as PCV) (%)
S1	Heroin	14.8	2.6	5.3	3
S2	Heroin	29.3	1.8	3.2	3
S3	Heroin	65.3	1.2	2.3	3

<sup>a</sup> Calculated from the assigned value.

<sup>b</sup> Robust between-laboratory CV with outliers removed, if applicable.


Of 102 results for which  $z$ -scores were calculated, 86 (84%) returned a  $z$ -score with  $|z| \leq 2.0$ , indicating an acceptable performance.

Twenty-four participants: **1, 2, 3, 6, 9, 10, 11, 12, 14, 15, 16, 17, 18, 19, 20, 23, 25, 26, 27, 28, 29, 30, 33** and **34** returned acceptable  $z$ -scores for all three samples. Ten participants returned at least one questionable or unacceptable  $z$ -score.

Laboratory **7**'s results were similar across all three samples, resulting in a very high  $z$ -score for Sample S1 and a very low  $z$ -score for Sample S3. This participant should check that they have reported results for the correct sample.

Laboratory **8** returned unacceptable  $z$ -scores across all reported results, with all results being higher than the assigned value (positive bias). This participant may have reported results as % salt (m/m) instead of % base (m/m) as requested for this PT study; otherwise, this participant should check their methodology for the cause of this positive bias.

The dispersal of participants'  $z$ -scores is presented graphically in Figure 5.



$z$ -Scores greater than 10.0 or less than -10.0 have been plotted at 10.0 or -10.0 respectively.

Figure 5  $z$ -Score Dispersal by Laboratory

#### 6.4 $E_n$ -Score


$E_n$ -Scores can be interpreted in conjunction with  $z$ -scores, as an unsatisfactory  $E_n$ -score can be caused by an inappropriate measurement, or uncertainty, or both. If a participant did not report an uncertainty with a result, an expanded uncertainty of zero (0) was used to calculate the  $E_n$ -score.

Of 102 results for which  $E_n$ -scores were calculated, 93 (91%) returned an acceptable  $E_n$ -score of  $|E_n| \leq 1.0$ , indicating agreement of the participant's result with the assigned value within their respective expanded uncertainties.

Thirty participants: **1, 3, 4, 5, 6, 8, 9, 10, 11, 12, 13, 14, 15, 16, 17, 18, 19, 20, 21, 22, 23, 25, 26, 27, 28, 29, 30, 32, 33** and **34** returned acceptable  $E_n$ -scores for all three samples. Four participants returned at least one unacceptable  $E_n$ -score.

Laboratories **2** and **24** returned unacceptable  $E_n$ -scores across all reported results; these participants did not report any uncertainties.

The dispersal of participants'  $E_n$ -scores is presented graphically in Figure 6.



$E_n$ -Scores greater than 10.0 or less than -10.0 have been plotted at 10.0 or -10.0 respectively.

Figure 6  $E_n$ -Score Dispersal by Laboratory

### 6.5 Identification of Cutting Agents

Cutting agents were added to each sample: niacinamide for Sample S1, sucrose for Sample S2, and procaine hydrochloride for Sample S3.

All participants reported on the identity of at least one sample's cutting agent. Results reported by participants are presented in Table 8.

Laboratories **3, 4, 7, 9, 10, 16, 17, 22, 25, 29, 30** and **31** correctly reported all cutting agents.

For Sample S1, all participants except Laboratory **21** reported on the identity of the cutting agent; all participants who did report for the cutting agent were correct.

For Sample S2, Laboratories **3, 4, 7, 9, 10, 16, 17, 22, 25, 29, 30** and **31** correctly identified sucrose as the cutting agent. Laboratory **13** identified 'sugars' however did not specify what type of sugar was present. Twenty-one participants did not report on the identity of the cutting agent.

For Sample S3, all participants except Laboratory **19** reported on the identity of the cutting agent; all those who did report for the cutting agent were correct.

## 6.6 Participants' Analytical Methods

Participants were requested to analyse the samples using their normal test methods and to report a single result for each sample as they would normally report to a client. Results reported in this way reflect the true variability of results reported to laboratory clients. The methodologies provided by participants are presented in Table 1.

A summary of accreditation status, participants' methods and reference standards is presented in Table 10.

Table 10 Summary of Participants' Analytical Methods

		Lab. Code
Accreditation	Yes to ISO/IEC 17025	1, 4, 6, 7, 8, 9, 11, 12, 13, 14, 15, 16, 18, 19, 20, 21, 22, 23, 25, 26, 27, 28, 29, 30, 31, 33, 34
	Not Accredited / Not Reported	2, 3, 5, 10, 17, 24, 32
Average Sample Mass Used per Analysis (mg)	< 20	3, 9, 13, 14, 23, 24, 29, 31, 32
	20 – 30	1, 4, 5, 6, 8, 10, 12, 17, 18, 19, 20, 22, 25, 26, 28, 30, 33, 34
	31 – 50	2, 11, 15, 21, 27
	51 - 100	7
	> 100	16
Conversion to Base?	Yes	1, 3, 6, 12, 13, 16, 18, 19, 20, 22, 23, 24, 25, 27, 32
	No	2, 4, 5, 7, 8, 9, 11, 14, 15, 17, 26, 28, 29, 30, 31, 33, 34
	Not Reported	10, 21
Instrument Used for Quantification	HPLC-DAD	4, 5, 10, 11, 17, 18, 24, 28, 31
	HPLC-UV/Vis	3, 9, 14, 25
	UPLC-DAD	2, 15, 26, 27, 33, 34
	UPLC-MS/MS	21
	GC-FID	1, 6, 7, 12, 13, 16, 19, 20, 22, 23, 29, 30, 32
	GC-MS	8
Solvent	Acetonitrile	7, 25
	Acetonitrile/Water	3, 9, 10, 14, 15, 31
	Acetonitrile/Water/Acid	4, 17
	Chloroform	8, 12, 16, 19, 20
	Chloroform/Methanol	1, 6
	Ethanol	2, 22, 29, 32
	Methanol	5, 11, 18, 23, 24, 27, 28, 30, 33, 34
	Other	13, 21, 26
Source of Calibration Standard	NMI Australia	2, 7, 10, 15, 19, 25, 26, 29
	Lipomed	5, 8, 11, 13, 14, 17, 21, 22, 23, 27, 31, 33, 34
	LGC	4, 24, 30
	British Pharmacopoeia	3, 9
	Other	1, 6, 12, 16, 18, 20, 28, 32

Plots of z-scores against various parameters are presented in Figures 7 to 11 (outliers have not been plotted). One participant used GC-MS for analysis, and they returned unacceptable z-scores across all samples (positive bias).




Figure 7 z-Score vs Sample Mass Used per Analysis




Figure 8 z-Score vs Sample Processing




Figure 9 z-Score vs Extraction Solvent




Figure 10 z-Score vs Measurement Instrument




Figure 11 z-Score vs Calibration Standard Source

### 6.7 Comparison of Results and Date of Analysis

As there were delays with sample delivery to some participants, the test samples were analysed over the course of approximately three months. There was no evidence of sample degradation over this period (Figure 12).




Figure 12 z-Score vs Sample Analysis Date

## 6.8 Comparison with Previous Heroin PT Studies

To enable direct comparison with previous NMI heroin PT studies, the target SD used to calculate  $z$ -scores has been kept constant at 3% PCV.

A summary of the satisfactory performance, presented as a percentage of the total number of scores, obtained by participants from 2014 to 2023 (last ten studies) is presented in Figure 13. The proportion of satisfactory  $z$ -scores and  $E_n$ -scores over this period on average is 84% for both.




Figure 13 Summary of Participants' Performance in NMI Heroin PT Studies

A number of participants have consistently participated in NMI heroin PT studies, and individual performance history reports are emailed to each participant at the end of each study. The consideration of  $z$ -scores for an analyte over time provides much more useful information than a single  $z$ -score. Over time, laboratories should expect at least 95% of their scores to lie within the range  $|z| \leq 2.0$ . Scores in the range  $2.0 < |z| < 3.0$  can occasionally occur, however, these should be interpreted in conjunction with the other scores obtained by that laboratory. For example, a trend of  $z$ -scores on one side of the zero line is an indication of method or laboratory bias.

A summary of individual laboratories' performances over the last ten studies is presented in Figures 14 and 15 for Australian and international laboratories respectively.  $z$ -Scores greater than 10.0 or less than -10.0 have been plotted at 10.0 or -10.0 respectively. Two Australian and four international laboratories have achieved acceptable  $z$ -scores across all samples in all heroin PT studies participated in over this period.




Figure 14 Summary of Australian Participants' z-Scores in NMI Heroin PT Studies




Figure 15 Summary of International Participants' z-Scores in NMI Heroin PT Studies



A comparison of all results from Australian and international laboratories in NMI heroin PT studies over the last ten studies is presented in Figure 16. Overall, both groups have performed very similarly, with Australian and international laboratories both achieving 84% acceptable z-scores respectively over this period.




Figure 16 Comparison of Australian and International Laboratories in NMI Heroin PT Studies

## 7 REFERENCES

Please note that for all undated references, the latest edition of the referenced document (including any amendments) applies.

- [1] ISO/IEC 17043:2010, *Conformity assessment – General requirements for proficiency testing*.
- [2] NMI, 2023, *Study Protocol for Proficiency Testing*, viewed December 2023, <[https://www.industry.gov.au/sites/default/files/2020-10/cpt\\_study\\_protocol.pdf](https://www.industry.gov.au/sites/default/files/2020-10/cpt_study_protocol.pdf)>.
- [3] NMI, 2023, *Chemical Proficiency Testing Statistical Manual*, viewed December 2023, <[https://www.industry.gov.au/sites/default/files/2019-07/cpt\\_statistical\\_manual.pdf](https://www.industry.gov.au/sites/default/files/2019-07/cpt_statistical_manual.pdf)>.
- [4] Thompson, M., Ellison, S.L.R. and Wood, R., 2006, ‘The International Harmonized Protocol for the Proficiency Testing of Analytical Chemistry Laboratories’, *Pure Appl. Chem.*, vol. 78, pp. 145-196.
- [5] ISO 13528, *Statistical methods for use in proficiency testing by interlaboratory comparison*.
- [6] Thompson, M., 2000, ‘Recent trends in inter-laboratory precision at ppb and sub-ppb concentrations in relation to fitness for purpose criteria in proficiency testing’, *Analyst*, vol. 125, pp. 385-386.
- [7] ISO/IEC 17025, *General requirements for the competence of testing and calibration laboratories*.
- [8] Eurachem/CITAC Guide CG 4, QUAM:2012.P1, *Quantifying Uncertainty in Analytical Measurement*, 3<sup>rd</sup> ed., viewed December 2023, <[http://www.eurachem.org/images/stories/Guides/pdf/QUAM2012\\_P1.pdf](http://www.eurachem.org/images/stories/Guides/pdf/QUAM2012_P1.pdf)>.
- [9] NATA, 2020, *Update to Measurement Uncertainty resources*, viewed December 2023, <<https://nata.com.au/news/update-to-measurement-uncertainty-resources/>>

## APPENDIX 1 ROBUST AVERAGE AND ASSOCIATED UNCERTAINTY, z-SCORE AND $E_n$ -SCORE CALCULATIONS

### A1.1 Robust Average and Associated Uncertainty

Robust averages were calculated using the procedure described in ISO 13528.<sup>5</sup> The associated uncertainties were estimated as according to Equation 4.

$$u_{rob\ av} = \frac{1.25 \times S_{rob\ av}}{\sqrt{p}} \quad \text{Equation 4}$$

where:

$u_{rob\ av}$  is the standard uncertainty of the robust average

$S_{rob\ av}$  is the standard deviation of the robust average

$p$  is the number of results

The expanded uncertainty ( $U_{rob\ av}$ ) is the standard uncertainty multiplied by a coverage factor of 2 at approximately 95% confidence level.

A worked example is set out below in Table 11.

Table 11 Uncertainty Estimate for Robust Average of Sample S2

Number of Results ( $p$ )	34
Robust Average	29.3% base (m/m)
$S_{rob\ average}$	0.95% base (m/m)
$u_{rob\ average}$	0.20% base (m/m)
$k$	2
$U_{rob\ average}$	0.40% base (m/m)

Therefore, the robust average for Sample S2 is  $29.3 \pm 0.4\%$  base (m/m).

### A1.2 z-Score and $E_n$ -Score Calculations

For each participant's result, a z-score and  $E_n$ -score are calculated according to Equations 2 and 3 respectively (Section 4).

A worked example is set out below in Table 12.

Table 12 z-Score and  $E_n$ -Score Calculation for Sample S1 Result Reported by Laboratory 1

Participant Result (% base (m/m))	Assigned Value (% base (m/m))	Target Standard Deviation	z-Score	$E_n$ -Score
$14.6 \pm 1.5$	$14.8 \pm 0.3$	3% as PCV, or: $0.03 \times 14.8 = 0.444\%$ base (m/m)	$z\text{-Score} = \frac{14.6-14.8}{0.444}$ = -0.45	$E_n\text{-Score} = \frac{14.6-14.8}{\sqrt{1.5^2+0.3^2}}$ = -0.13

## APPENDIX 2 ACRONYMS AND ABBREVIATIONS

ASCLD/LAB	American Society of Crime Laboratory Directors – Laboratory Accreditation Board
CITAC	Cooperation on International Traceability in Analytical Chemistry
CRM	Certified Reference Material
CV	Coefficient of Variation
DAD	Diode Array Detection
EA	European Accreditation
FID	Flame Ionisation Detection
GAG	General Accreditation Guidance (NATA)
GC	Gas Chromatography
GUM	Guide to the expression of Uncertainty in Measurement
HPLC	High Performance Liquid Chromatography
IEC	International Electrotechnical Commission
ISO	International Organization for Standardization
Max	Maximum
Md	Median
Min	Minimum
MS	Mass Spectrometry
MS/MS	Tandem Mass Spectrometry
MU	Measurement Uncertainty
N	Number of numeric results
NATA	National Association of Testing Authorities, Australia
NMI	National Measurement Institute, Australia
NR	Not Reported
PCV	Performance Coefficient of Variation
PDA	Photodiode Array Detection
PT	Proficiency Testing
RA	Robust Average
RM	Reference Material
SD	Standard Deviation
SI	International System of Units
SS	Spiked Samples
UPLC	Ultra Performance Liquid Chromatography
UV/Vis	Ultraviolet/Visible Detection

**END OF REPORT**