



# NATIONAL STANDARDS COMMISSION

G. H.  
6/4C/150  
18/8/89

## NATIONAL MEASUREMENT (PATTERNS OF INSTRUMENTS) REGULATIONS

### REGULATION 9

#### CERTIFICATE OF APPROVAL No 6/4C/150

This is to certify that an approval for use for trade has been granted in respect of the pattern and variant of the

Avery Model 1104 Weighing Instrument.

#### CONDITIONS OF APPROVAL

##### Special:

This approval does NOT permit the verification of new instruments of the pattern. It does not become due for review.

This approval applies only to instruments in the field which have been previously verified; they may be re-verified, or verified after repairs, at the discretion of the relevant Weights and Measures authority.

Instruments purporting to comply with this approval shall be marked NSC No 6/4C/150 prior to re-verification.

Instruments may be relocated and reinstalled but shall not be altered in any way other than to replace damaged components with equivalent components.

Signed

Executive Director

#### Descriptive Advice

Pattern: approved 15/2/89

- Avery model 1104 self-indicating weighing instrument of 1 kg capacity with a verification scale interval of 0.010 kg.

Variant: approved 15/2/89

1. Of 500 g capacity with a verification scale interval of 5 g.

#### Test Procedure

Instruments should be tested in accordance with the relevant sections of the Inspector's Handbook. The results shall not exceed the maximum permissible errors given in Document 118, 2nd Edition, October 1986.

#### Filing Advice

The documentation for this approval comprises only this Certificate.