



**CANCELLED**

COMMONWEALTH OF AUSTRALIA

**NATIONAL STANDARDS COMMISSION**

**Weights and Measures  
(National Standards)  
Act 1960-1966**

**Weights and Measures  
(Patterns of Instruments)  
Regulations**

# *Certificate of Approval*

---

**CERTIFICATE NUMBER 6/9C/17**

*In respect of the pattern of*

Avery 3202 CLE Platform Weighing Machine and Variant 1.

Submitted and  
manufactured by: W. & T. Avery (Aust.) Pty. Ltd.,  
3-5 Birmingham Avenue,  
Villawood,  
New South Wales. 2163.

This is to certify that the pattern and variant of the instrument described in this Certificate have been examined by the National Standards Commission under the provisions of the abovementioned Regulations and have been approved as being suitable for use for trade.

Approval was granted for the pattern and variant 1 on 3rd September, 1968.

Approval was granted on condition that all instruments made in conformity with the pattern and variants:

1. are appropriately marked NSC No 6/9C/17 and, where required by State legislation, with the State approval number also; and
2. comply with the General Specifications for Weighing and Measuring Instruments to be Used for Trade, in

23/9/70

Cont'd over

respect of that part of the instrument which was not previously approved by a State.

This Certificate comprises:

Pages 1 to 3 dated 23rd September, 1970.

Pursuant to regulation 12 of the abovementioned Regulations, this Certificate is applicable in the State of Victoria.

Date of issue 23rd September, 1970.

Signed

A handwritten signature in cursive script, appearing to read "Robert A. Hamilton".

A person authorised by the Commission to sign Certificates under the abovementioned Regulations.

---

DESCRIPTION OF PATTERN

The pattern is of a self-indicating platform weighing machine of 135 lb capacity, 110 lb by 4 oz graduations on the dial and 25 lb by 4 oz graduations on the tare bar.

The pattern is a modification of an instrument previously approved in the State of Victoria under Notice of Approval No 311, which has been modified by replacing the counting headwork with a self-indicating dial headwork incorporating a CLA spring-resistant mechanism as approved in Victorian Notice of Approval No 201.

DESCRIPTION OF VARIANT

1. Having other capacities up to a maximum of 320 lb or 150 kg.

GENERAL NOTES

Notice of approval of the pattern and variant described in this Certificate was given in Memorandum of Approval No 133 dated 9th September, 1968.



# NATIONAL STANDARDS COMMISSION

CERTIFICATE OF APPROVAL No 6/9C/17

VARIATION No 1

This is to certify that the pattern and variants of the

Avery Self-indicating Platform Weighing Machine Model 3202 CLE

approved under regulation 12 in Certificate No 6/9C/17 dated 23  
September 1970,

have been approved without territorial limitations under the Weights and  
Measures (Patterns of Instruments) Regulations as being suitable for  
use for trade when converted to read in metric units in accordance with  
Appendix 8 of the General Specifications for Measuring Instruments to  
be Used for Trade.

Date of Approval: 16 September 1974

This approval is subject to review on or after 31 December 1975 to  
determine the extent to which the pattern and variants may need to be  
modified to comply with the Commission's pattern approval requirements  
at the date of review.

All instruments conforming to this approval shall be marked  
"NSC No 6/9C/17".

Signed

Executive Officer