

# CANCELLED



## NATIONAL STANDARDS COMMISSION

CERTIFICATE OF APPROVAL No 6/9C/23

Approved: Avery 3205 CLM Platform Weighing Instrument

This is to certify that the pattern of the above instrument has been approved under the Weights and Measures (Patterns of Instruments) Regulations as being suitable for use for trade when converted to read in metric units with a dial capacity of 100 kg  $\times$  0.2 kg and a tare capacity of 35 kg.

Date of Approval: 4 February 1974

Description of Pattern:

The pattern may be identified by reference to the following approvals:

Queensland	...	Approval No 661
South Australia	...	Approval No 44
Tasmania	...	Approval dated 9 December 1964

Approval has been granted for 600 graduations on the dial (as approved in Queensland).

An instrument converted to read in metric units is shown in Figure 1.

Conditions of Approval:

1. This approval expires on 31 January 1976.
2. All instruments conforming to the pattern or variants shall be marked "NSC No 6/9C/23".

Signed

A person authorised by the Commission to sign Certificates under the abovementioned Regulations.

8/2/74

FIGURE 6/9C/23 - 1



Avery 3205 CLM Platform Weighing Instrument

8/2/74