



Australian Government

National Measurement
Institute

Bradfield Road, West Lindfield NSW 2070

Interim Provisional Supplementary Certificate of Approval NMI PS659

VALID FOR VERIFICATION PURPOSES UNTIL 21 AUGUST 2014

Issued by the Chief Metrologist under Regulation 60
of the
National Measurement Regulations 1999

This is to certify that an approval for use for trade has been granted in respect of the instruments herein described.

METTLER TOLEDO Model DataBridge™ Point of Sale (POS) System

submitted by Mettler-Toledo Limited
220 Turner Street
Port Melbourne VIC 3207

NOTE: This Certificate relates to the suitability of the pattern of the instrument for use for trade only in respect of its metrological characteristics. This Certificate does not constitute or imply any guarantee of compliance by the manufacturer or any other person with any requirements regarding safety.

DOCUMENT HISTORY

Rev	Reason/Details	Date
0	Pattern provisionally approved – interim certificate issued	26/02/14

CONDITIONS OF APPROVAL

General

Instruments purporting to comply with this approval shall be marked with pattern approval number 'NMI PS659' and only by persons authorised by the submittor.

It is the submittor's responsibility to ensure that all instruments marked with this approval number are constructed as described in the documentation lodged with the National Measurement Institute (NMI) and with the relevant Certificate of Approval and Technical Schedule. Failure to comply with this Condition may attract penalties under Section 19B of the National Measurement Act and may result in cancellation or withdrawal of the approval, in accordance with document NMI P 106.

Special Conditions of Approval: (weighbridges)

The pattern has not been assessed for compliance with requirements which are outside the scope of document NMI M7, including those features which control the automation of weighbridge operation or ticket formats for public weighbridges.

This Certificate does not constitute or imply approval for these functions. Details of these requirements can be found on the NMI website.

Special Conditions of Approval: (Provisional Approval)

The locations or serial numbers of instruments may be obtained from the National Measurement Institute. The submitter shall advise NMI in writing of the proposed location or serial number of each instrument prior to it being initially verified.

Instruments purporting to comply with this approval shall be marked with approval number 'NMI PS659' and only by persons authorised by the submitter. (Note: The 'P' in the approval number may be a temporary marking.)

The approval will remain provisional pending completion of satisfactory testing and evaluation.

In the event of unsatisfactory performance the approval may be cancelled (or altered).

The submitter shall implement such modifications as required by NMI. In the event that such modifications (if any are required by NMI) are not made to the satisfaction of NMI, this approval may be withdrawn.

1. Description of Pattern provisionally approved on 26/02/14

A METTLER TOLEDO model DataBridge™ system to provide certain additional facilities for transactions when interfaced to compatible (#) NMI-approved measuring instruments granted with reference to document NMI M7.

1.1 Key Features

- The system provides point of sale arrangements when connected to NMI-approved measuring instruments fitted with a METTLER TOLEDO model IND780 digital indicator (approval NMI S502) or other compatible (#) NMI-approved measuring instruments.
 - The system receives measurement data from the output interface of the approved measuring instrument and computes prices using a product look up (PLU) facility.
 - The system computes total price for multiple items including non-measured items and is approved for use for transactions direct to the public.
 - Manually entered measurement data shall be indicated as such on a printed transaction record.
 - The system is able to apply a tare value up to the maximum capacity of the approved measuring instrument. Preset tare values may be keyboard-entered or stored (e.g. within a PLU facility).
 - The POS controllers may be connected in a network to share common PLU data, to accumulate and retrieve management information including information pertaining to pricing, material codes, vendor details, etc.
- (#) 'Compatible' is defined to mean that no additions/changes to the hardware/software specified in this approval are required for satisfactory operation of the system.

1.2 System Description

The METTLER TOLEDO model DataBridge™ system comprises:

(i) POS Controller

The METTLER TOLEDO model DataBridge™ POS controller comprises a DELL model Optiplex 9010 computer or equivalent (*) PC-based device that operates a Microsoft Windows-based operating system running MTC2.d11 with checksum version '1839567634' and DataBridgeLRC.d11 with checksum version '469821645' legally relevant software modules. Non-legally relevant component of the software is also shown as a second version number. Both these numbers may be displayed by clicking on the 'i' (information button) on the DataBridge™ screen on the POS controller.

- (*) 'Equivalent' is defined to mean other proprietary equipment of the same or better specifications requiring no changes to the software specified in this approval for satisfactory operation of the system.

(ii) Electronic Indications

Indications shall satisfy the requirements of document NMI M7, *Pattern Approval Specifications for Point of Sale Systems*.

A Dell computer monitor or equivalent (*) is connected to the POS controller to provide an indication for the operator and the customer.

(iii) Printing Devices

Transaction records shall satisfy the requirements of document NMI M7, *Pattern Approval Specifications for Point of Sale Systems*.

A Hewlet Packard model HP1000 printer or equivalent (*) is connected to the controller to provide transaction record printing facility.

Note: Tickets have NOT been assessed for compliance with the requirements for Weighbridge Measurement Tickets as given in relevant Licensing Directives of the trade measurement section of NMI as published on the NMI website.

- (*) 'Equivalent' is defined to mean other proprietary equipment of the same or better specifications requiring no changes to the software specified in this approval for satisfactory operation of the system.

(iv) Multiple Instruments Facility

The METTLER TOLEDO model DataBridge™ POS system may be connected to up to 6 approved measuring instruments. The POS system is configured to display which measuring instrument is connected.

The measuring instrument to be used is preselected by the operator when they log into the METTLER TOLEDO model DataBridge™ application.

Note: In the case of this feature, each instrument/combo shall be clearly identified to correspond to the appropriate measuring instrument display shown on the POS system display. Trade measurement authorities may require additional markings or signs to ensure that these relationships are clear.

(v) Truck Weighing Functions

Providing functions intended specifically for truck weighing applications, including provision for 'truck and product' identification data to be stored in memory.

The truck weighing functions provide for:

- simple vehicle weighing, where the gross weight of a vehicle is determined by a single weighing;
- first/second weighing, where a vehicle is weighed before and after a loading or unloading operation;
- function keys programmed to perform various functions (such as accessing and searching stored vehicle, item, product or client information).

(vi) Additional System Facilities

The system may include additional peripheral devices including but not limited to barcode scanning devices, RFID card readers, driver control stations, programmable logic controllers (PLC), input/output controllers, video surveillance cameras, video overlay devices and other plant/site-specific control systems. The facilities shall not interact with the system in a way that would cause an incorrect indication of the measured quantity or price.

2. Description of Variant 1 provisionally approved on 26/02/14

With the METTLER TOLEDO model DataBridge™ software module and intended to be used for weighing operations using small platform weighing instruments used to weigh small amounts of scrap material.

The METTLER TOLEDO model DataBridge™ can be interfaced with up to a maximum of 6 weighing instruments.

INTERIM TEST PROCEDURE No PS659

The POS system shall be tested for compliance with the requirements of General Supplementary Certificate No S1/0/A dated 20/03/92, as follows:

Testing requires a minimum of 5 checks spanning the measurement range of the measuring instrument. For each check, ensure the device is correctly:

- repeating the result of the primary indicator; and/or
- summing several primary indicators; and/or
- printing the results.

Ensure the printed format and the display format is in compliance with General Supplementary Certificate No S1/0/A.

Perform a display segment check for indicators with this function.

For network systems, check that the measurement data printed on the transaction record is correctly reproduced from each device connected in the network.

Signed by a person authorised by the Chief Metrologist to exercise their powers under Regulation 60 of the *National Measurement Regulations 1999*.



Dr A Rawlinson