

National Measurement Institute

12 Lyonpark Road, North Ryde NSW 2113

Cancellation Supplementary Certificate of Approval No S290

Issued by the Chief Metrologist under Regulation 60 of the
National Measurement Regulations 1999

This is to certify that the approval for use for trade granted in respect of the

Revere Transducers Model BSP-A3-500-30P5 Load Cell

submitted by Thermo Ramsey Pty Ltd

(formerly Ramsey Technology)

now of 11 Box Road

Taren Point NSW 2229

has been cancelled in respect of new instruments as from 1 June 2005.

Signed by a person authorised by the Chief Metrologist to exercise his powers under Regulation 60 of the National Measurement Regulations 1999.



Supplementary Certificate of Approval

No S290

Issued under Regulation 9
of the
National Measurement (Patterns of Measuring Instruments) Regulations

This is to certify that an approval for use for trade has been granted in respect of the

Revere Transducers Model BSP-A3-500-30P5 Load Cell

submitted by Ramsey Technology

20-22 Box Road

Taren Point NSW 2229.

NOTE: This Certificate relates to the suitability of the pattern of the instrument for use for trade only in respect of its metrological characteristics. This Certificate does not constitute or imply any guarantee of compliance by the manufacturer or any other person with any requirements regarding safety.

CONDITIONS OF APPROVAL

This approval is subject to review on or after 1 October 1997. This approval expires in respect of new instruments on October 1998.

Instruments purporting to comply with this approval shall be marked NSC No S290 and only by persons authorised by the submittor.

Instruments incorporating a component purporting to comply with this approval shall be marked NSC No S290 in addition to the approval number of the instrument.

Page 2

The Commission reserves the right to examine any instrument or component of an instrument purporting to comply with this approval.

It is the submittor's responsibility to ensure that all instruments marked with this approval number are constructed as described in the documentation lodged with the Commission and with the relevant Certificate of Approval and Technical Schedule. Failure to comply with this Condition may attract penalties under Section 19B of the National Measurement Act and may result in cancellation or withdrawal of the approval, in accordance with the Commission's Document 106.

The values of the performance criteria (maximum number of scale intervals etc.) applicable to an instrument incorporating the pattern approved herein shall be within the limits specified herein and in any approval documentation for the other components.

DESCRIPTIVE ADVICE

Pattern:

approved 21 September 1992

A Revere Transducers model BSP-A3-500-30P5 load cell of 227 kg capacity.

Variant:

approved 21 September 1992

1. Model BSP-A3-250-30P5 of 120 kg capacity.

Technical Schedule No S290 describes the pattern and variant 1.

Variant:

approved 24 December 1993

2. Model BSP-A3-2.5K-30P5 of 1134 kg capacity.

Technical Schedule No S290 Variation No 1 describes variant 2.

Variant:

approved 12 April 1994

3. Model BSP-A3-1K-30P5 of 450 kg capacity.

Technical Schedule No S290 Variation No 2 describes variant 3.

FILING ADVICE

Supplementary Certificate of Approval No S290 dated 14 March 1994 is superseded by this Certificate and may be destroyed.

The documentation for this approval now comprises:

Supplementary Certificate of Approval No S290 dated 8 July 1994 Technical Schedule No S290 dated 17 November 1992 (incl. Table 1) Technical Schedule No S290 Variation No 1 dated 14 March 1994 (incl. Table 2)

Technical Schedule No S290 Variation No 2 dated 8 July 1994 (incl. Table 3 and Notification of Change)

Figures 1 and 2 dated 17 November 1992

Signed and sealed by a person authorised under Regulation 9 of the National Measurement (Patterns of Measuring Instruments) Regulations to exercise the powers and functions of the Commission under this Regulation.

J. Birch



TECHNICAL SCHEDULE No S290

Pattern:

A Revere Transducers Model BSP-A3-500-30P5 Load Cell.

Submittor:

Ramsey Engineering Pty Ltd

20-22 Box Road

Taren Point NSW 2229.

1. Description of Pattern

A Revere Transducers model BSP-A3-500-30P5 load cell of 227 kg capacity (refer Figure 1 and Table 1) approved for use with a maximum of 500 verification scale intervals.

1.1 Method of Mounting

Mounting is to be in accordance with the manufacturer's instructions and as shown in Figure 2.

1.2 Markings

The following is the minimum data required to be marked on the load cells:

Manufacturer's name or mark Model number Serial number NSC approval number Maximum rated capacity

NSC No S290

2. Description of Variant 1

Model BSP-A3-250-30P5 of 120 kg capacity approved for use with a maximum of 1500 verification scale intervals (refer Table 1).

TABLE 1

Type: Revere Transducers BSP-A3-###-30P5

Maximum capacity	120 kg	227 kg
Maximum number of verification	J	3
scale intervals	1500	500
Minimum value of verification		
scale interval	0.031 kg	0.035 kg
Minimum dead load output return for	_	_
multi-range instruments (DR)	0.031 kg	0.059 kg
Output rating (nominal)	3 (m\	
Input impedance (nominal)	350 (oh	
Supply voltage (AC or DC)	10 (V)	10 (V)
Cable length (± 0.1 m)	5 (m)	5 (m)
Number of leads (plus shield)	4	4

NOTE: ### - '500' for the pattern (227 kg) and '250' for variant 1 (120 kg).



TECHNICAL SCHEDULE No S290

VARIATION No 1

Pattern:

A Revere Transducers Model BSP-A3-500-30P5 Load Cell.

Submittor:

Ramsey Technology 20-22 Box Road

Taren Point NSW 2229.

1. Description of Variant 2

A Revere Transducers model BSP-A3-2.5K-30P5 load cell of 1134 kg capacity (refer Figure 1 and Table 2) approved for use with a maximum of 2000 verification scale intervals.

1.1 Method of Mounting

Mounting is to be in accordance with the manufacturer's instructions and as shown in Figure 2.

TABLE 1

Type: Revere Transducers BSP-A3-2.5K-30P5

Maximum capacity	1134 kg
Maximum number of verification	
scale intervals	2000
Minimum value of verification	
scale interval	0.15 kg
Minimum dead load output return for	J
multi-range instruments (DR)	0.245 kg
Output rating (nominal)	3 mV/V
Input impedance (nominal)	350 ohms
Supply voltage (AC or DC)	10 V
Cable length (± 0.1 m)	4 m
Number of leads (plus shield)	4
·	



TECHNICAL SCHEDULE No S290

VARIATION No 2

Pattern:

A Revere Transducers Model BSP-A3-500-30P5 Load Cell.

Submittor:

Ramsey Technology

20-22 Box Road

Taren Point NSW 2229.

1. Description of Variant 3

A Revere Transducers model BSP-A3-1K-30P5 load cell of 450 kg capacity (refer Figure 1 and Table 3) approved for use with a maximum of 1000 verification scale intervals.

1.1 Method of Mounting

Mounting is to be in accordance with the manufacturer's instructions and as shown in Figure 2.

TABLE 3

Type: Revere Transducers BSP-A3-1K-30P5

Maximum capacity	450	kg
Maximum number of verification		Ü
scale intervals	1000	
Minimum value of verification		
scale interval	0.091	ka
Minimum dead load output return for		3
multi-range instruments (DR)	0.0675	kg
Output rating (nominal)		mV/V
Input impedance (nominal)	350	ohms
Supply voltage (AC or DC)	10	V
Cable length (± 0.1 m)	4	m
Number of leads (plus shield)	4	

NOTIFICATION OF CHANGE

In Technical Schedule No S290 Variation No 1 dated 14 March 1994 the caption for the table should be amended to read 'Table 2'.



National Standards Commission Notification of Change Supplementary Certificate of Approval No S290 Change No 1

The following changes are made to the approval documentation for the

Revere Transducers Model BSP-A3-500-30P5 Load Cell

submitted by Ramsey Technology

20-22 Box Road

Taren Point NSW 2229.


In Supplementary Certificate of Approval No S290 dated 8 July 1994;


- (i) The Condition of Approval referring to the review of the approval should be amended to read:
 - "This approval becomes subject to review on or after 1 October 2002."
- (ii) The Condition of Approval referring to the expiry of the approval should be deleted.

Signed and sealed by a person authorised under Regulation 9 of the National Measurement (Patterns of Measuring Instruments) Regulations to exercise the powers and functions of the Commission under this Regulation.

J. Bunh

FIGURE S290 - 1





Typical Mounting Method