




REFERENCE MATERIAL PRODUCT INFORMATION SHEET

NMIA D981: α -Naphyrone hydrochloride

Report ID: D981.2014.04

Chemical Formula: C₁₉H₂₃NO.HCl

Molecular Weight: 317.9 g/mol (HCl), 281.4 g/mol (base)



Property value

Batch No.	CAS No.	Purity estimate
12-D-09	1349245-31-3	89.9 ± 3.1%

IUPAC name: 1-(1-Naphthyl)-2-(1-pyrrolidinyl)-1-pentanone hydrochloride (1:1).

Expiration of certification: The property values are valid till 24 July 2017, i.e. three years from the date of re-certification provided the **unopened** material is handled and stored in accordance with the recommendations below. The material as issued in the unopened container and stored as recommended below should be suitable for use beyond this date, subject to confirmation of batch stability from the issuing body. The expiry date/shelf life does not apply to sample bottles that have been opened. In such cases it is recommended that the end-user conduct their own in-house stability trials. The material will be re-tested on an annual basis to ensure that the property values are still valid. In the event a product fails the stability trial, notification will be sent to all impacted customers.

Description: Off-white solid sourced from external supplier, and certified for identity and purity by NMIA. Packaged in amber glass bottles with a septum and crimped aluminium cap or screw top cap.

Intended use: This reference material is recommended for qualitative analysis only.

Instructions for use: Equilibrate the bottled material to room temperature before opening.

Recommended storage: When not in use this material should be stored at or below 25 °C in a closed container in a dry, dark area.

Stability: In the absence of long term stability data the stability of this material has been judged from stability trials conducted on similar materials by NMI Australia over the last ten years. The long-term stability of the compound in solution has not been examined.

Homogeneity assessment: The homogeneity of the material was assessed using purity assay by GC-FID on ten randomly selected 1-2 mg sub samples of the material. The material was judged to be sufficiently homogeneous at this level of sampling as the variation in analysis results between samples was not significantly different at a 95% confidence level from that observed on repeat analysis of the same sample.

Safety: Treat as a hazardous substance. Use appropriate work practices when handling to avoid skin or eye contact, ingestion or inhalation of dust. Refer to the provided safety data sheet.

S. R. Davies

Dr Stephen R. Davies,
Team Leader,
Chemical Reference Materials, NMI.
20 September 2022

This report supersedes any issued prior to 20 September 2022.

NATA Accreditation No. 198 / Corporate Site No. 14214.

Legal notice: Terms and Conditions associated with the provision of this reference material can be found on the NMIA website.

Characterisation Report:

The identity was confirmed by a range of spectroscopic techniques, NMR, IR and MS. Impurities of related structure were assessed by GC-FID. The purity estimate was obtained by mass balance from a combination of traditional analytical techniques, including GC-FID, thermogravimetric analysis, Karl Fischer analysis and ¹H NMR spectroscopy. The purity value is calculated as per Equation 1.

$$\text{Purity} = (100\% - I_{\text{ORG}}) \times (100\% - I_{\text{VOL}} - I_{\text{NVR}}) \quad \text{Equation 1}$$

I_{ORG} = Organic impurities of related structure, I_{VOL} = volatile impurities, I_{NVR} = non-volatile residue.

Supporting evidence is provided by qualitative headspace GC-MS analysis of occluded solvents and elemental microanalysis.

GC-FID: Instrument: Varian CP-3800
 Column: VF-1MS or HP-1, 30 m \times 0.32 mm I.D. \times 0.25 μ m
 Program: 150 $^{\circ}$ C (1 min), 10 $^{\circ}$ C/min to 250 $^{\circ}$ C (1 min), 30 $^{\circ}$ C/min to 300 $^{\circ}$ C (3 min)
 Injector: 250 $^{\circ}$ C
 Detector Temp: 320 $^{\circ}$ C
 Carrier: Helium
 Split ratio: 20/1
 Relative peak area response of main component:
 Initial analysis: Mean = 97.3%, s = 0.2% (10 sub samples in duplicate, August 2012)
 Initial analysis: Mean = 97.4%, s = 0.2% (10 sub samples in duplicate, August 2012) (HP-1)

GC-FID: Instrument: Varian CP-3800 or Agilent 6890N
 Column: HP-5 or HP-1, 30 m \times 0.32 mm I.D. \times 0.25 μ m
 Program: 150 $^{\circ}$ C (1 min), 10 $^{\circ}$ C/min to 240 $^{\circ}$ C (2 min), 20 $^{\circ}$ C/min to 300 $^{\circ}$ C (3 min)
 Injector: 250 $^{\circ}$ C
 Detector Temp: 320 $^{\circ}$ C
 Carrier: Helium
 Split ratio: 20/1
 Relative peak area response of main component:
 Initial analysis: Mean = 97.3%, s = 0.04% (8 sub samples in duplicate, August 2013)
 Re-analysis: Mean = 97.1%, s = 0.10% (5 sub samples in duplicate, July 2014)

Karl Fischer analysis: Moisture content 5.8% mass fraction (August 2012)
 Moisture content 6.9% mass fraction (July 2013)
 Moisture content 7.0% mass fraction (July 2014)

Thermogravimetric analysis: Non-volatile residue < 0.2% mass fraction (September 2011). The volatile content (e.g. organic solvents and/or water) could not be determined because of the inherent volatility of the material.

Spectroscopic and other characterisation data

GC-MS:	Instrument:	Agilent 6890/5973
	Column:	TG-1MS, 30 m x 0.25 mm I.D. x 0.25 μ m
	Program:	150 °C (1 min), 10 °C/min to 250 °C (5 min), 15 °C/min to 300 °C (3 min)
	Injector:	250 °C
	Split ratio:	20/1
	Transfer line temp:	300 °C
	Carrier:	Helium, 1.0 mL/min
	Scan range:	50-550 <i>m/z</i>
	The retention time of the free base is reported along with the major peaks in the mass spectrum. The latter are reported as mass/charge ratios and (in brackets) as a percentage relative to the base peak.	
	Free base (11.2 min):	280 (8), 279 (34), 236 (17), 208 (12), 181 (13), 155 (16), 127 (69), 126 <i>m/z</i>
LC/ESI -MS:	Instrument:	Waters Acquity UPLC/TQD
	Operation:	Positive ion mode, direct infusion at 5 μ L/min
	Ionisation:	ESI spray voltage at 3.0 kV positive ion
	EM voltage:	650 kV
	Cone voltage:	20 V
	Peak:	282.2 (M+H ⁺) <i>m/z</i>
HS-GC-MS:	Instrument:	Agilent 6890/5973/G1888
	Column:	DB-624, 30 m x 0.25 mm I.D. x 1.4 μ m
	Program:	50 °C (5 min), 7 °C/min to 120 °C, 15 °C/min to 220 °C (8.3 min)
	Injector:	150 °C
	Split ratio:	50/1
	Transfer line temp:	280 °C
	Carrier:	Helium, 1.2 mL/min
	Solvents detected:	Ethanol, diethyl ether, butanal
TLC:	Conditions:	Kieselgel 60F ₂₅₄ . CH ₂ Cl ₂ /methanol (9:1) Single spot observed, R _f = 0.49. Visualisation with UV at 254 nm
IR:	Instrument:	Biorad FTS3000MX FT-IR
	Range:	4000-400 cm ⁻¹ , KBr powder
	Peaks:	3454, 3386, 2965, 2874, 2653, 2475, 1688, 1573, 1510, 1445, 1335, 1238, 1095, 1073, 935, 792 cm ⁻¹
¹ H NMR:	Instrument:	Bruker Avance DMX-600
	Field strength:	600 MHz
	Solvent:	D ₂ O (4.79 ppm)
	Spectral data:	δ 0.66 (3H, t, <i>J</i> = 7.9 Hz), 1.06-1.24 (2H, m), 1.90-2.26 (6H, m), 3.14 (1H, m), 3.33 (1H, m), 3.76 (2H, m), 5.26 (1H, dd, <i>J</i> = 3.9, 7.1 Hz), 7.59-7.63 (2H, m), 7.68 (1H, ddd, <i>J</i> = 1.3, 6.9, 6.9 Hz), 7.98 (1H, d, <i>J</i> = 8.0 Hz), 8.06 (1H, dd, <i>J</i> = 0.7, 7.3 Hz), 8.17 (1H, d, <i>J</i> = 8.3 Hz), 8.44 (1H, d, <i>J</i> = 8.4 Hz) ppm Ethanol, diethyl ether and butanal estimated at 3.7%, 0.7 and ~ 0.0% respectively mass fraction were observed in the ¹ H NMR (August 2012) Ethanol, diethyl ether and butanal estimated at 0.41%, 0.23% and 0.04% respectively mass fraction were observed in the ¹ H NMR (August 2013)
¹³ C NMR:	Instrument:	Bruker Avance DMX-600
	Field strength:	150 MHz
	Solvent:	D ₂ O (referenced to the CH ₂ of residual ethanol in D ₂ O: 58.05 ppm)
	Spectral data:	δ 13.8, 17.8, 23.4, 23.5, 32.5, 52.6, 55.8, 71.6, 124.9, 125.3, 127.9, 129.7, 129.8, 130.1, 130.7, 131.9, 134.4, 136.0, 200.7 ppm
Melting point:	117-119 °C	
Microanalysis:	Found:	C = 67.3%; H = 8.0%; N = 4.0%; Cl = 10.3% (August, 2012)
	Calculated:	C = 71.8%; H = 7.6%; N = 4.4%; Cl = 11.2% (Calculated for C ₁₉ H ₂₃ NO.HCl)