




# REFERENCE MATERIAL PRODUCT INFORMATION SHEET

## NMIA D1006: $\alpha$ -Dimethylaminopentiphenone hydrochloride

Report ID: D1006.2016.05

Chemical Formula:  $C_{13}H_{19}NO.HCl$

Molecular Weight: 241.8 g/mol (HCl), 205.3 g/mol (base)



### Property value

Batch No.	CAS No.	Purity estimate
13-D-23	2748622-76-4	99.2 $\pm$ 2.8%

**Synonyms:**  $\alpha$ -Dimethylaminovalerophenone hydrochloride,  
1-Phenyl-2-(dimethylamino)pentan-1-one hydrochloride

**Expiration of certification:** The property values are valid till 8 August 2019, i.e. three years from the date of re-certification provided the **unopened** material is handled and stored in accordance with the recommendations below. The material as issued in the unopened container and stored as recommended below should be suitable for use beyond this date, subject to confirmation of batch stability from the issuing body. The expiry date/shelf life does not apply to sample bottles that have been opened. In such cases it is recommended that the end-user conduct their own in-house stability trials.

**Description:** White powder prepared by synthesis, and certified for identity and purity by NMIA. Packaged in amber glass bottles with a septum and crimped aluminium cap or screw top cap.

**Intended use:** This reference material is recommended for qualitative analysis only.

**Instructions for use:** Equilibrate the bottled material to room temperature before opening.

**Recommended storage:** When not in use this material should be stored at or below 25 °C in a closed container in a dry, dark area.

**Stability:** This material has demonstrated stability over a minimum period of three years. The long-term stability of the compound in solution has not been examined.

**Homogeneity assessment:** The homogeneity of the material was assessed using purity assay by GC-FID on ten randomly selected 1-2 mg sub samples of the material. The material was judged to be sufficiently homogeneous at this level of sampling as the variation in analysis results between samples was not significantly different at a 95% confidence level from that observed on repeat analysis of the same sample.

**Safety:** Treat as a hazardous substance. Use appropriate work practices when handling to avoid skin or eye contact, ingestion or inhalation of dust. Refer to the provided safety data sheet.

S. R. Davies

Dr Stephen R. Davies,  
Team Leader,  
Chemical Reference Materials, NMI.  
17 November 2023

This report supersedes any issued prior to 17 November 2023.

NATA Accreditation No. 198 / Corporate Site No. 14214.

**Legal notice:** Terms and Conditions associated with the provision of this reference material can be found on the NMIA website.

---

### Characterisation Report:

The identity was confirmed by a range of spectroscopic techniques, NMR, IR and MS. Impurities of related structure were assessed by GC-FID. The purity value was obtained by mass balance from a combination of traditional analytical techniques, including GC-FID, thermogravimetric analysis, Karl Fischer analysis and <sup>1</sup>H NMR spectroscopy. The purity value is calculated as per Equation 1.

$$\text{Purity} = (100 \% - I_{\text{ORG}}) \times (100 \% - I_{\text{VOL}} - I_{\text{NVR}}) \quad \text{Equation 1}$$

$I_{\text{ORG}}$  = Organic impurities of related structure,  $I_{\text{VOL}}$  = volatile impurities,  $I_{\text{NVR}}$  = non-volatile residue.

Supporting evidence is provided by headspace GC-MS analysis of occluded solvent, elemental microanalysis and quantitative nuclear magnetic resonance (qNMR). The purity estimate by qNMR was obtained using the two proton multiplet at 1.97 ppm against a certified internal standard of maleic acid.

**Note: This material exists as a mixture of hydrochloride and hydrobromide salts in approximately 4:1 molar ratio. The stated purity value represents the combined mass fraction of these salts.**

**Based on qNMR analysis 1 mg of D1006 is equivalent to 812.8 ± 24 µg of  $\alpha$ - dimethylaminopentiophenone free base.**

GC-FID:	Instrument:	Varian CP-3800
	Column:	VF-1ms or HP-5, 30 m × 0.32 mm I.D. × 0.25 µm
	Program:	120 °C (1 min), 10 °C/min to 160 °C (2 min), 30 °C/min to 300 °C (3 min)
	Injector:	250 °C
	Detector Temp:	320 °C
	Carrier:	Helium
	Split ratio:	20/1
	Relative peak area of the main component as the free base:	
	Initial analysis:	Mean = 99.3%, s = 0.07% (10 sub samples in duplicate, October 2013) [VF-1]
	Re-analysis:	Mean = 99.4%, s = 0.03% (5 sub samples in duplicate, September 2014) [VF-1]
	Initial analysis:	Mean = 99.3%, s = 0.07% (10 sub samples in duplicate, October 2013) [HP-5]
	Re-analysis:	Mean = 99.5%, s = 0.04% (7 sub samples in duplicate, August 2015) [HP-5]
	Re-analysis:	Mean = 99.7%, s = 0.01% (5 sub samples in duplicate, August 2016) [HP-5]
Karl Fischer analysis:	Moisture content 0.1% mass fraction (November 2013)	
	Moisture content < 0.1% mass fraction (2014, 2015 and August 2016)	
Thermogravimetric analysis:	The volatile content (e.g. organic solvents and/or water) could not be determined because of the inherent volatility of the material and/or degradation at elevated temperatures. The non volatile residue < 0.2% mass fraction (October 2013)	
QNMR:	Instrument:	Bruker DMX-600
	Field strength:	600 MHz
	Solvent:	D <sub>2</sub> O
	Internal standard:	Maleic acid (98.7% mass fraction)
	Initial analysis:	Mean (1.97 ppm) = 99.2%, s = 0.1% (1 sub sample, February 2012)

## Spectroscopic and other characterisation data

GC-MS:	Parent compound:	
	Instrument:	Agilent 6890/5973
	Column:	TG-1MS, 30 m x 0.25 mm I.D. x 0.25 $\mu$ m
	Program:	60 °C (1 min), 10 °C/min to 300 °C (3 min)
	Injector:	250 °C,
	Split ratio:	20/1
	Transfer line temp:	280 °C
	Carrier:	Helium, 1.0 mL/min
	Scan range:	50-550 <i>m/z</i>
	The retention time of the parent compound is reported with the major peaks in the mass spectra. The latter are reported as mass/charge ratios and (in brackets) as a percentage relative to the base peak.	
	Free base (12.5 min):	205 ( $M^+$ , 1), 203 (4), 162 (8), 105 (29), 100 (100), 84 (15), 77 (55), 71 (47), 58 (52), 51 (20), 42 (32) <i>m/z</i>
LC/ESI-MS:	Instrument:	Micromass Quatro LC Micro
	Operation:	Positive ion mode, direct infusion at 10 $\mu$ L/min
	Ionisation:	ESI spray voltage at 3.0 kV positive ion
	EM voltage:	650 V
	Cone voltage:	15 V
	Peak:	206.3 ( $M+H^+$ ) <i>m/z</i>
HS-GC-MS:	Instrument:	Agilent 6890/5973/G1888
	Column:	DB-624, 30 m x 0.25 mm I.D. x 1.4 $\mu$ m
	Program:	50 °C (5 min), 7 °C/min to 120 °C, 15 °C/min to 220 °C (8.3 min)
	Injector:	150 °C
	Split ratio:	50/1
	Transfer line temp:	280 °C
	Carrier:	Helium, 1.2 mL/min
	Solvents detected:	Ethanol, propanol, butanol and diethyl ether
TLC:	Conditions:	Kieselgel 60F <sub>254</sub> . Dichloromethane/methanol/triethylamine (96/10/2) Single spot observed, $R_f = 0.72$ . Visualisation with UV at 254 nm
IR:	Instrument:	Biorad FTS3000MX FT-IR
	Range:	4000-400 $cm^{-1}$ , KBr powder
	Peaks:	3343, 3056, 2962, 2939, 2716, 2666, 2492, 1679, 1596, 1580, 1470, 1454, 1350, 1234, 1001, 929, 718 $cm^{-1}$
<sup>1</sup> H NMR:	Instrument:	Bruker Avance III-400
	Field strength:	400 MHz
	Solvent:	D <sub>2</sub> O (4.79 ppm)
	Spectral data:	$\delta$ 0.81 (3H, t, $J = 7.3$ Hz), 1.18 (2H, m), 2.07 (2H, m), 2.99 (6H, s), 5.26 (1H, dd, $J = 1.4, 4.6$ Hz), 7.64 (2H, dd, $J = 1.7, 7.5$ Hz), 7.80 (1H, tt, $J = 1.2, 7.5$ Hz), 8.05 (2H, dd, $J = 1.3, 8.4$ Hz), ppm Ethanol estimated at 0.11% mass fraction was observed in the <sup>1</sup> H NMR ppm
<sup>13</sup> C NMR:	Instrument:	Bruker Avance III-400
	Field strength:	100 MHz
	Solvent:	D <sub>2</sub> O
	Spectral data:	$\delta$ 12.9, 16.7, 30.4, 41.0 (br), 69.7, 128.9, 129.4, 133.6, 135.8, 197.3 ppm
Melting point:		176-181 °C
Microanalysis:	Found:	C = 62.1%; H = 8.1%; N = 5.6%; Cl% = 11.2% (December 2013)
	Calculated:	C = 62.3%; H = 8.0%; N = 5.6%; Cl% = 11.3% [Calculated for (C <sub>13</sub> H <sub>19</sub> NO.HCl) <sub>4</sub> . (C <sub>13</sub> H <sub>19</sub> NO.HBr)]