

#### National Measurement Institute

# Appointment as a Verifying Authority for Reference Standards of Measurement

In accordance with Regulation 73 of *National Measurement Regulations 1999* (Cth), in force under the *National Measurement Act 1960* (Cth), the Chief Metrologist hereby appoints,

## Cl Scientific Pty Limited (ABN 35 105 113 595)

### Operating at: Unit 1, 9 Bearing Road Seven Hills NSW 2147

to be a Verifying Authority for the verification of reference standards of measurement under regulation 13 of the *National Measurement Regulations 1999* (Cth) for the following physical quantities:

### **Mass and Density**

This appointment is for the period from 29 October 2024 to 28 October 2027 and is limited to the range specified in the attached schedule, and the use of procedures approved by the Chief Metrologist.

Dated this Nineteenth day of November 2024

Signed

James Cantrill For Dr Richard Bruce Warrington Chief Metrologist National Measurement Institute

Certificate: NMI2024-025-CIS-VARSM Form No. NMI/VARSM/2024

Page 1 of 4

## Schedule to Appointment as a Verifying Authority for Reference Standards of Measurement

| <b>CI Scientific Pty Limite</b> | be |
|---------------------------------|----|
| (ABN 35 105 113 595             | )  |

Operating at: Unit 1, 9 Bearing Road Seven Hills NSW 2147

| Physical Quantity | Range of Standard  | Least Uncertainty   |
|-------------------|--|---|
| Mass              | from 1 mg to 1000 kg   |   |
| Mass standards    | from 1 mg to 20 mg<br>above 20 mg up to 50 mg<br>above 50 mg up to 500 mg<br>above 500 mg up to 1 g<br>above 1 g up to 2 g<br>above 2 g up to 5 g<br>above 2 g up to 5 g<br>above 20 g up to 20 g<br>above 20 g up to 50 g<br>above 50 g up to 100 g<br>above 100 g up to 200 g<br>above 200 g up to 500 g<br>above 500 g up to 1 kg<br>above 1 kg up to 2 kg<br>above 2 kg up to 10 kg<br>above 2 kg up to 10 kg<br>above 20 kg up to 20 kg | 3 μg<br>4 μg<br>5 μg<br>6 μg<br>7 μg<br>10 μg<br>25 μg<br>30 μg<br>40 μg<br>60 μg<br>0.11 mg<br>0.3 mg<br>0.6 mg<br>2.5 mg<br>20 mg<br>0.2 g<br>0.02 kg |
|                   | above 250 kg up to 500 kg<br>above 500 kg up to 1000 kg  | 0.05 kg<br>0.1 kg   |

James Cantrill For Dr Richard Bruce Warrington Chief Metrologist National Measurement Institute

Certificate: NMI2024-025-CIS-VARSM Form No. NMI/VARSM/2024

Page 2 of 4

#### Physical Quantity

#### Range of Standard

#### Least Uncertainty

#### Density

| Density Hydrometers          | from 0 to 2.0 g/mL       | 1/3 of smallest scale division |
|------------------------------|--------------------------|--------------------------------|
| Apparent Density Hydrometers | from 0.990 to 1.040 kg/L | 1/3 of smallest scale division |
| Soil Hydrometers             | from -5 to +60 g/L       | 1/3 of smallest scale division |
| Brix Hydrometers             | from 0 to 90 Bx          | 1/3 of smallest scale division |
| Bark Hydrometers             | from 0 to 80 Bk          | 1/3 of smallest scale division |
| Baume Hydrometers            | from 0 to 90 Be          | 1/3 of smallest scale division |
| Specific Gravity Hydrometers | from 0.6 to 2.0          | 1/3 of smallest scale division |
| Alcohol Hydrometers          | from 0 to 100 %          | 1/3 of smallest scale division |

#### Signatories

The following persons are the permitted signatories under this appointment:

| Name            | Physical Quantity | Range                             |
|-----------------|-------------------|-----------------------------------|
| Michael Cyrulla | Mass and Density  | as per the scope of this schedule |
| Jurgen Cyrulla  | Mass and Density  | as per the scope of this schedule |
| Hardik Mavani   | Mass and Density  | as per the scope of this schedule |

#### **Statutory Conditions**

This appointment as a verifying authority for reference standards of measurement under regulation 73 of the *National Measurement Regulations 1999* (Cth) is subject to the conditions stated in regulation 77 of the *National Measurement Regulations 1999* (Cth) as amended. At the time of appointment regulation 77 contains the following conditions

- (a) That the authority participate in training, related to the performance of the duties of an authority, required by the Chief Metrologist;
- (b) That the authority report, as required by the Chief Metrologist, about its performance of its duties;

James Cantrill For Dr Richard Bruce Warrington Chief Metrologist National Measurement Institute

Certificate: NMI2024-025-CIS-VARSM Form No. NMI/VARSM/2024

Page 3 of 4

- (c) That the authority, and any responsible agent or employee of the authority, comply with the *National Measurement Act 1960* (Cth) and the *National Measurement Regulations 1999* (Cth) and any condition stated in the instrument of appointment.
- (d) That the authority comply with any determinations applying to the authority under regulation 20 of the *National Measurement Regulations 1999* (Cth).

#### **Additional Conditions**

In addition to the statutory conditions of appointment of authorities contained in regulation 77 of the *National Measurement Regulations 1999* (Cth) this appointment is also subject to the following conditions:

- (i) Continuing accreditation against AS ISO/IEC 17025 *General requirements for the competence of testing and calibration laboratories* in the form of NATA accreditation No. 411
- (ii) The authority shall not engage a responsible agent or arrange for any standard of measurement to be verified by an agent or anyone under its supervision without obtaining the prior consent of the Chief Metrologist in writing;
- (iii) Discharge of all financial obligations to the Chief Metrologist and/or the National Measurement Institute in respect of this appointment;
- (iv) Compliance with the formatting and/or any other requirements of the Chief Metrologist and/or the National Measurement Institute with respect to certificates of verification of reference standards of measurement;
- (v) This appointment revokes and replaces any previous appointments and/or any extensions granted to any previous appointments.

Notes: Nil

James Cantrill For Dr Richard Bruce Warrington Chief Metrologist National Measurement Institute

Certificate: NMI2024-025-CIS-VARSM Form No. NMI/VARSM/2024